



Federal Office
for Information Security

Assurance Continuity Maintenance Report

BSI-DSZ-CC-0519-V6-2025-MA-01

**Stationäres eHealth Kartenterminal ORGA 6141
online, Version 3.9.2:1.2.0 / 3.9.2:2.0.0**

from

Worldline Healthcare GmbH



SOGIS
Recognition Agreement
for components up to
EAL 4

The IT product identified in this report was assessed according to the procedures on Assurance Continuity [1] and the developer's Impact Analysis Report (IAR). The baseline for this assessment was the Certification Report, the Security Target and the Evaluation Technical Report of the product certified by the Federal Office for Information Security (BSI) under BSI-DSZ-CC-0519-V6-2025.

The certified product itself did not change. The changes are related to an update of the user guidance.

Considering the nature of the change leads to the conclusion that it is classified as a minor change and that certificate maintenance is the correct path to continuity of assurance.

The resistance to attacks has not been re-assessed in the course of this maintenance process. Therefore, the assurance statement as outlined in the Certification Report BSI-DSZ-CC-0519-V6-2025 dated 16.10.2025 is of relevance and has to be considered when using the product. Details can be found on the following pages.

This report is an addendum to the Certification Report BSI-DSZ-CC-0519-V6-2025.



Common Criteria
Recognition Arrangement
recognition for components
up to EAL 2 and ALC_FLR
only



Bonn, 30 October 2025

The Federal Office for Information Security

Bundesamt für Sicherheit in der Informationstechnik

Godesberger Allee 87 - D-53175 Bonn - Postfach 20 03 63 - D-53133 Bonn
Phone +49 228 99 9582-0 - Fax +49 228 9582-5477 - Infoline +49 228 99 9582-111

Assessment

The IT product identified in this report was assessed according to the procedures on Assurance Continuity [1] and the Impact Analysis Report (IAR) [2]. The baseline for this assessment was the Certification Report of the certified product (Target of Evaluation, TOE) [3], its Security Target and the Evaluation Technical Report as outlined in [3].

The vendor for the Stationäres eHealth Kartenterminal ORGA 6141 online, Version 3.9.2:1.2.0 / 3.9.2:2.0.0, Worldline Healthcare GmbH, submitted an IAR [2] to the BSI for approval. The IAR is intended to satisfy the requirements according to the procedures on Assurance Continuity [1]. In accordance with those requirements, the IAR describes (i) the changes made to the certified TOE, (ii) the evidence updated as a result of the changes and (iii) the security impact of the changes.

The certified product itself did not change. The Security Target was editorially updated.

The changes are related to an update of the user guidance [5] only. A clarification for the usage of the remote management interface for initial pairing is added.

Conclusion

The maintained change is at the level of guidance documentation. The change has no effect on product assurance, but the updated guidance documentation has to be followed.

Considering the nature of the change leads to the conclusion that it is classified as a minor change and that certificate maintenance is the correct path to continuity of assurance.

The resistance to attacks has not been re-assessed in the course of this maintenance process. Therefore, the assurance statement as outlined in the Certification Report BSI-DSZ-CC-0519-V6-2025 dated 16.10.2025 is of relevance and has to be considered when using the product.

Obligations and notes for the usage of the product:

All aspects of assumptions, threats and policies as outlined in the Security Target not covered by the TOE itself need to be fulfilled by the operational environment of the TOE.

The customer or user of the product shall consider the results of the certification within his system risk management process. In order for the evolution of attack methods and techniques to be covered, he should define the period of time until a re-assessment for the TOE is required and thus requested from the sponsor of the certificate.

Additional Note: The strength of the cryptographic algorithms was not rated in the course of the product certification and this maintenance procedure (see BSIG¹ Section 9, Para. 4, Clause 2).

- Use of the latest version of the TOE guidance documentaion, see [5].

¹ Act on the Federal Office for Information Security (BSI-Gesetz - BSIG) of 14 August 2009, Bundesgesetzblatt I p. 2821

This report is an addendum to the Certification Report [3].

References

- [1] Common Criteria document “Assurance Continuity: CCRA Requirements”, version 3.1, 29 February 2024
Common Criteria document “Assurance Continuity: SOG-IS Requirements”, version 1.2, March 2024
- [2] IAR, V 0.1.0, 28.10.2025, Worldline Healthcare GmbH (confidential document)
- [3] Certification Report BSI-DSZ-CC-0519-V6-2025 for Stationäres eHealth Kartenterminal ORGA 6141 online Version 3.9.2:1.2.0 / 3.9.2:2.0.0, 16.10.2025, Bundesamt für Sicherheit in der Informationstechnik
- [4] Security Target for the Evaluation of the Product ORGA 6141 online, Version 4.1.27, 29.10.2025, Worldline Healthcare GmbH
- [5] Bedienungsanleitung Stationäres eHealth-Kartenterminal ORGA 6141 online Hardware-Version 1.2.0 und 2.0.0 (ORGA Neo) mit Firmware-Version 3.9.2, Version 25.9.2, Oktober 2025, Worldline Healthcare GmbH